

Certificate

Product Name : Switching Power Supply
Brand Name : MEAN WELL
Model Number : SHP-30K-xy
(x=55, 115, 230, 380, 400CP, y= Blank,-PM, -MOD)
Applicant : MEAN WELL ENTERPRISES CO., LTD.
Address : No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248,
Taiwan (R.O.C.)
Report Number : S2C32-M020-2305-441
Issue Date : May 28, 2025
Applicable Standards :

Emission :

EN 55032:2015/A11:2020
EN 61000-3-12:2011
EN IEC 61000-3-11:2019

Immunity:

EN 55035:2017/A11:2020
EN IEC 61000-6-2:2019
EN 61000-4-2:2009
EN IEC 61000-4-3:2020
EN 61000-4-4:2012
EN 61000-4-5:2014/A1:2017
EN 61000-4-6:2014
EN 61000-4-8:2010
EN 61000-4-34:2007/A1:2009



Based on the EMC Directive 2014/30/EU and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



Central Research Technology Co.

11, Lane 41, Fushuen St., Jungshan Chiu,
Taipei 104, Taiwan
Tel : 886-2-25872719
Fax : 886-2-25850364


(Jim Chien/Authorized Signatory)

Date: May 28, 2025

TAF 0905

FCC CAB Code TW1104, TW0019

NVLAP Lab Code 200575-0

ISED CAB Code TW0905

VCCI Accep. No. R-11527, C-11609, T-11441, G-10010, C-20010

T-20009, G-10614